

Member State
P.R.China

OIML Certificate No.
R76/2006-CN1-16.01



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

Applicant

Name: Mettler Toledo (ChangZhou) Measurement Technology Ltd.

Address: No. 111, west Taihu Road, Xinbei District, Changzhou, Jiangsu. (213125) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Terminal, as part of a non-automatic weighing instrument

Type	ICS429 ICS439 ICS449 ICS469 ICS489
Accuracy class	III
Maximum number of verification scale interval n	100000
Fraction of the maximum permissible error (p_i)	0
weighing range	Single interval / Multi-interval / Multiple range
Power supply	AC 100-240 V, 50/60 Hz
Software identification	XY-XX-01.YY.YY-XX-YY (X is a character from A to Z; Y is a number between 0 to 9.)
List of devices:	semi-automatic zero-setting device, automatic zero-setting device, initial zero-setting device, tare balancing or tare-weighing device.

The numbers of the associated test reports is NIM-LSmc-1601 including 15 pages, NIM-LSmc-1602 including 15 pages and NIM-LSmc-1603 including 29 pages.

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 2006

For accuracy class (if applicable): Ⅲ

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority

Office of OIML Affairs

中国秘书处

October 30, 2016

The OIML member

吴清海

October 30, 2016

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.