

Member State of OIML
Germany



OIML Certificate N°
R76/2006-DE1-09.01
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Dirk Ratschko

Applicant

Name: Sartorius AG
Address: Weender Landstr. 94-108, 37075 Göttingen

Manufacturer of the certified type is the applicant.

Identification of the certified type Nonautomatic electromechanical weighing instrument

Type: MSX

Further characteristics see tables 1, 2 and 3 in pages 2 and 3

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006

for accuracy classes **I** and **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

OIML Certificate N°
R76/2006-DE1-09.01
Revision 1

This revision 1 was issued because the new weighing module types SD EE, SE EA, SF EA, higher weighing range type SC EA, motorized levelling, motorized wind protector and ionizer were added. The conformity was established by tests described in the associated Report N° 1.12-4041251, Revision 1, and the associated Test Reports N° 1.12-4041251/10, N° 1.12-4041251/11, N° 1.12-4041251/12, N° 1.12-4041251/13, N° 1.12-4041251/14 and N° 1.12-4041251/15. The test results of the former test reports N° 1.12-4041251/1, N° 1.12-4041251/2, N° 1.12-4041251/3, N° 1.12-4041251/4, N° 1.12-4041251/5, N° 1.12-4041251/6, N° 1.12-4041251/7, N° 1.12-4041251/8 and N° 1.12-4041251/9 remain valid.

The Issuing Authority



Dr. D. Ratschko
 Oberregierungsrat

14.06.2010



The OIML Member



Dr. R. Schwartz
 Direktor und Professor

14.06.2010

The weighing instruments are modular structured, consisting of weighing module and indicating and operator terminal. The modules have in addition a type designation of their own, e.g. indicating and operator terminal: YAC01MSE..., YAC01MSU... YAC01MSA...
 weighing module: S... E... (SA EA, SA EB,....)

Table 1

Module type	SA EA	SB EA	SC EA
Class	(I)		
Max	50 g ... 320 g	500 g ... 3200 g	5000 g ... 14200 g
e =	1 mg ... 5 mg	10 mg ... 50 mg	100 mg ... 200 mg
d =	0,1 mg ... 5 mg	1 mg ... 50 mg	10 mg ... 200 mg
n ≤	320000	320000	142000
Tare-balancing range ≤	100 % of Max	100 % of Max	100 % of Max
Temperature range	+ 15 °C / + 25 °C	+ 15 °C / + 25 °C	+ 15 °C / + 25 °C
Temperature range ¹⁾	+ 5 °C / + 40 °C	+ 5 °C / + 40 °C	+ 5 °C / + 40 °C
Nominal load of load receptor ₂₎	362 g	3400 g	14620 g
Initial zero-setting range + dead load ≤	312 g	2900 g	9620 g

Table 2

Module type	SD EE	SE EA	SF EA
Class	(I)		
Max	50 g ... 220 g	50 g ... 520 g	1000 g ... 5200 g
e =	1 mg ... 2 mg	1 mg ... 5 mg	10 mg ... 50 mg
d =	0,01 mg ... 2 mg	0,1 mg ... 5 mg	1 mg ... 50 mg
n ≤	220000	520000	520000
Tare-balancing range ≤	100 % of Max	100 % of Max	100 % of Max
Temperature range	+ 15 °C / + 25 °C	+ 15 °C / + 25 °C	+ 15 °C / + 25 °C
Temperature range ¹⁾	+ 5 °C / + 40 °C	+ 5 °C / + 40 °C	+ 5 °C / + 40 °C
Nominal load of load receptor ₂₎	242 g	572 g	5400 g
Initial zero-setting range + dead load ≤	192 g	522 g	4400 g

Table 3

Module type	SB EB	SC EB	SC EA	SC EC
Class	(II)			
Max	1 g ... 620 g	500 g ... 6200 g	500 g ... 8200 g	5000 g ... 12200 g
e =	0,01 g ... 0,1 g	0,1 g ... 0,5 g	0,1 g ... 1 g	1 g ... 2 g
d =	0,001 g ... 0,1 g	0,01 g ... 0,5 g	0,01 g ... 1 g	0,1 g ... 2 g
n ≤	62000	62000	82000	12200
Tare-balancing range ≤	100 % of Max	100 % of Max	100 % of Max	100 % of Max
Temperature range	+ 10 °C / + 30 °C	+ 10 °C / + 30 °C	+ 10 °C / + 30 °C	+ 10 °C / + 30 °C
Temperature range ¹⁾	+ 5 °C / + 40 °C	+ 5 °C / + 40 °C	+ 5 °C / + 40 °C	+ 5 °C / + 40 °C
Nominal load of load receptor ₂₎	682 g	6820 g	8620 g	13420 g
Initial zero-setting range + dead load ≤	681 g	6320 g	8120 g	8420 g

¹⁾ Only for weighing instruments with incorporated span adjustment device with automatic release.

²⁾ The sum of Max, initial zero setting range and dead load shall not exceed the nominal load of the load receptor.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.