# Physikalisch-Technische Bundesanstalt

### Braunschweig und Berlin

Member State of OIML Germany



OIML Certificate N° **R76/2006-DE1-09.02** 

## OIML CERTIFICATE OF CONFORMITY

### **Issuing Authority**

Name:	Physikalisch-Technische Bundesanstalt
Address:	Bundesallee 100, 38116 Braunschweig
Person responsible:	Dr. Panagiotis Zervos

### Applicant

Name:	Bizerba GmbH & Co. KG	
Address:	Wilhelm-Kraut-Str. 65, 72336 Balingen GERMANY	

Manufacturer of the certified type is the applicant.

Identification of the cer-	Nonautomatic electromechanical weighing instrument with
tified type	or without lever system
	Туре: WM

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, edition 2006 for accuracy classes (II) and (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

## Physikalisch-Technische Bundesanstalt

### OIML Certificate N° R76/2006-DE1-09.02

The conformity was established by tests described in the report N° 1.12-4043571 and the associated test reports No. 1.12-4043571/1 (Samples 1, WM, complete instrument), No. 1.12-4043571/2 (Sample 2, additional EMC-Tests) and No. 1.12-4043571/3 (Type WM as modul).

#### The Issuing Authority

#### The CIML Member

Dr. P. Zervos **Direktor und Professor** 

Dr. R. Schwartz Direktor und Professor

08.12.2009

08.12.2009

#### Table 1

Accuracy class		
Max	1 kg … 120 t	1 kg 120 t
$n \leq 1$	6000	1000
$n_i \leq 2^{(2)}$	3000	1000
Temperature range	-10 °C to 40 °C	-10 °C to +40 °C
Tare balancing range	100 % of Max	100 % of Max
Tare input range	100 % of Max <sup>1)</sup>	100 % of Max <sup>1)</sup>
	100 % of Max <sub>1</sub> <sup>2)</sup>	100 % of Max <sub>1</sub> <sup>2)</sup>

<sup>1)</sup> This applies to each range of single- and multiple range instruments <sup>2)</sup> This applies only to multi-interval instruments

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial guotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.