

Member State of OIML  
Germany



OIML Certificate No.  
**R76/2006-DE1-10.01**  
Revision 1

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. Oliver Mack

### Applicant

Name: Bizerba GmbH & Co. KG  
Address: Wilhelm-Kraut-Str. 65  
72336 Balingen  
Germany

### Manufacturer

Name:	Bizerba GmbH & Co. KG	Bizerba (Shanghai) WeighTech & Systems Co., Ltd.
Address:	Wilhelm-Kraut-Str. 65 72336 Balingen Germany	D-3 Qianpu Road, Shuhui Real Estate Park East New Area of Songjiang Industrial Zone, Shanghai, 201611 P.R. China

**Identification of the certified type** Non-automatic weighing instrument for direct sales to the public  
Type: SC II...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

OIML Certificate No.  
**R76/2006-DE1-10.01**  
**Revision 1**

With the 1<sup>st</sup> revision new verification scale intervals for the weighing systems WS36 were added. Furthermore a variant with 7-inch-display and the possibility of data transmission via wireless LAN interface module "WL" were added.

The conformity was established by the results of tests and examinations provided in the associated Test Reports No. 1.12-4044914/4 (42 pages), No. 1.12-4044914/5 (6 pages), No. 1.12-4044914/6 (41 pages), No. 1.12-4044914/7 (6 pages), 1.12-4044914/8 (14 pages), 1.12-4044914/9 (14 pages) and 1.12-4044914/10 (14 pages). The test results of the former test reports No. 1.12-4044914/1, No. 1.12-4044914/2 and No. 1.12-4044914/3 remain valid

**The Issuing Authority**



Dr. O. Mack  
 Member of certification body



**The OIML Member**



Dr. R. Schwartz  
 Head of Division

22.07.2013

22.07.2013

The weighing ranges with Max, Min, e, d and number of verification scale intervals may be chosen within the limits of No. 3.2 of R 76-1 and of Table 1.

Table 1

Accuracy class	III
Type	SC II...
Max ≤	300kg
e =	1 g ... 50 g
d =	1 g ... 50 g
n ≤	7500
Tare balancing range (subtractive)	≤ 100 % of Max
Preset tare range	≤ Max ≤ Max <sub>1</sub> at multi-interval instruments
Temperature range	-10 °C / 40 °C

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.