



**Physikalisch-Technische Bundesanstalt
Braunschweig und Berlin**
Nationales Metrologieinstitut

Member State of OIML
Germany



OIML Certificate No.
R76/2006-DE1-16.01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. O. Mack

Applicant

Name: Bizerba GmbH & Co. KG
Address: Wilhelm-Kraut-Str. 65, 72336 Balingen

Manufacturer of the certified type is the applicant.

Identification of the certified type

Non-automatic price-computing weighing instrument for direct sales to the public
Type: MC...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006
for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



OIML Certificate No.
R76/2006-DE1-16.01

The conformity was established by the results of tests and examinations provided in the associated Test Reports

No. 1.12-4073750/1 that includes 16 pages
No. 1509030104 that includes 56 pages

The Issuing Authority

The OIML Member

Dr. O. Mack
Member of Certification Body

Dr. R. Schwartz
Vice President

15.02.2016

15.02.2016

Identification of the pattern (continued)

Weighing instrument with load receptor, terminal with touch, keyboard and intern or extern printer.
The weighing ranges with Max, Min, e and number of verification scale intervals may be chosen within the limits of No. 3.2 of R76-1 and of Table 1.

Table 1

Accuracy class	III
Maximum capacity Max	30 kg
Verification scale interval e	1 g ... 10 g
Number of verification scale intervals	≤ 6000
Tare-balancing range (subtractive)	≤ 100 % • Max
Preset tare range	≤ 100 % • Max
Temperature range	-10 °C / +40 °C

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.