



Member State  
Denmark

OIML Certificate N°  
R76/2006-DK3-13.01

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: **DELTA**  
Address: Venlighedsvej 4  
2970 Hørsholm  
Denmark

Person responsible: J. Hovgård Jensen

### Applicant

Name: **Kosan Crisplant A/S**  
Address: P. O. Pedersens Vej 22  
8200 Aarhus N  
Denmark

### Manufacturer

of the certified pattern: **Kosan Crisplant A/S**

### Identification

of the certified pattern: **Non-automatic weighing instrument**  
**Type: CUC-Ex**  
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
**edition 2006**  
**for accuracy class III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages



## OIML Certificate N° R76/2006-DK3-13.01

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1911335, dated 08-07-2011 that includes 68 pages

The issuing authority: **DELTA, OIML Issuing Authority DK3**  
17 May 2013

J. Hovgård Jensen  
Certification Officer

### Characteristics

Accuracy class		III
Weighing range		Single-interval
Number of intervals	Max / e	3000
Minimum input voltage	$\Delta U_{\min}$	1.0 $\mu$ V per VSI
Measuring range	Min. / Max.	0 mV / 10 mV
Excitation voltage	$U_{\text{exc}}$	5 VDC
Load cell impedance	Min. / Max.	350 ohm / 1200 ohm
Load cell connecting system		4-wire, shielded
Module fractional factor	$p_i$	0.5 for the indicator
Interface		Protective, according to paragraph 5.3.6
Connected load cells		Shall comply with R60
Supply voltage		100 to 240 VAC (50/60 Hz), or 12 VDC battery supplied
Temperature range for the indicator		-10 °C / +55 °C
Software identification	CUC LCM	Version 3.xx.yy Version 1.01.zz where xx, yy, zz have values from 00 to 99 representing the non-legal software.

#### Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.