Member State Denmark

OIML Certificate N° R76/2006-DK3-14.06

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **DELTA**

Address: Venlighedsvej 4

2970 Hørsholm

Denmark

Person responsible: J. Hovgård Jensen

Applicant

Name: **Kaifeng Group Co., Ltd.**Address: No 6-8, Shifang Middle Road

Gushan Town, Yongkang City,

Zhejiang 321307

CHINA

Manufacturer

of the certified pattern: Kaifeng Group Co., Ltd.

Identification

of the certified pattern: Non-automatic weighing instrument

Type: ACS-C / ACS-L2 / ACS-L5 / BL-CN / ACS-R / BL-GN / ACS-A1 / ACS-268 / ACS-568 / ACS-668A /

ACS-668B / ACS-768 / ACS-868

Intended to be used for direct sales to the public. Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76 edition 2006 for accuracy class III

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages

OIML Certificate N° R76/2006-DK3-14.06

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1913671, dated 12-12-2013 that includes 68 pages, DELTA, DK, No. DANAK-1914393, dated 12-08-2014 that includes 20 pages.

The issuing authority: **DELTA, OIML Issuing Authority DK3**

21 October 2014

J. Hovgård Jensen Certification Officer

Characteristics

Accuracy class		III
Weighing range		Single-interval or dual-range
Maximum capacity	Max	3 kg to 30 kg
Minimum capacity	Min	20 e
Verification scale interval	e	≥ 1 g
Number of intervals	Max _i / e _i	≤ 3000
Subtractive tare	Т	≤-Max
Interface		None
Supply voltage		220 VAC, 50 Hz,
		optional internal battery
Temperature range for the indicator		-10 °C / +40 °C
Software identification:		120305

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.

Page 2. This certificate includes 2 pages