



Member State  
Denmark

OIML Certificate N°  
R76/2006-DK3-16.08

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: **DELTA**  
Address: Venlighedsvej 4  
2970 Hørsholm  
Denmark

Person responsible: J. Hovgård Jensen

### Applicant

Name: **Hao Han Industry Company.**  
Address: 1F, No. 8, Alley 10, Lane 114, Sec. 2, Xingnan Rd.,  
Zhonghe Dist.,  
New Taipei City 235  
Taiwan R.O.C

### Manufacturer

of the certified pattern: **Hao Han Industry Company.**

### Identification

of the certified pattern: **Non-automatic weighing instrument**  
**Type: ATxy**  
Further characteristics are set out on page 2 and 3.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
**edition 2006**  
**for accuracy class III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 3 pages



## OIML Certificate N° R76/2006-DK3-16.08

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1916850, dated 12-09-2016 that includes 60 pages

The issuing authority: **DELTA, OIML Issuing Authority DK3**  
9 November 2016

J. Hovgård Jensen  
Certification Manager

### Characteristics

Accuracy class		III
Weighing range		Single-interval or dual-range
Maximum capacity	Max <sub>i</sub>	1.5 kg to 30 kg
Minimum capacity	Min <sub>i</sub>	20 e
Verification scale interval	e <sub>i</sub>	≥ 0.5 g
Number of intervals	Max <sub>i</sub> / e <sub>i</sub>	≤ 6000 for single-interval, ≤ 3000 per range for dual-range
Subtractive tare	T	≤ -Max
Interface		Protective, according to paragraph 5.3.6
Supply voltage		9 VDC from external adapter supplied from 100-240 VAC, 50-60 Hz, 6 V internal battery (optional)
Temperature range for the indicator		-10 °C / +40 °C
Software identification:	Legal SW application SW other	See page 3

### Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.



## OIML Certificate N° R76/2006-DK3-16.08

The software consists of a model depending part with an identification and a common software with a version number.

The software identification is displayed during the power-up sequence of the instrument and shall have the following values.

MODEL NAME	APPLICATION	SOFTWARE IDENTIFICATION
AT5	WEIGHING	CE020
AT5K	CHECKWEIGHING	CE004
AT5P	PRICE COMPUTING	CE005
AT5C	PARTS COUNTING	CE006
AT6	WEIGHING	CE011
AT6K	CHECKWEIGHING	CE012
AT6P	PRICE COMPUTING	CE013
AT6C	PARTS COUNTING	CE014
AT7	WEIGHING	CE007
AT7K	CHECKWEIGHING	CE008
AT7P	PRICE COMPUTING	CE009
AT7C	PARTS COUNTING	CE010

The common software, which version can be seen using the setup command Fn1, has the following approved version: 1.02xx, where xx can be from 00 to 99 and covers differences in non-legal applications