

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-10.20

Revision 1

Project number 10201010

Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

Number NMI-10201010-01 dated 13 April 2011 that includes 60 pages;

Number NMI-10201010-02 dated 18 July 2011 that includes 25 pages.

Characteristics of the non-automatic weighing instrument:

Indicator, as a part of a non-automatic weighing instrument.

Type: DI-5000

The maximum number of verification scale intervals will be:

$n \leq 3000$ for class III instruments or;

$n \leq 1000$ for class IIII instruments.

Applied error fraction p_i : 0,5;

Minimum input impedance: 300 Ω ;

Maximum input impedance: 1050 Ω ;

Signal voltage per verification scale interval $\geq 1,67 \mu\text{V}$;

Temperature range $-10^\circ\text{C} / +40^\circ\text{C}$.