



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-12.06
Project number 12219005
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Mettler-Toledo (Albstadt) GmbH Unter dem Malesfelsen 34 D-72458 Albstadt Germany
Manufacturer	Mettler-Toledo
Identification of the certified type	A Non-automatic weighing instrument Type : ICS..
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **(II)**, **(III)** or **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
8 March 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-12.06
Project number 12219005
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- Number R76/2006-NL1-10.25 dated 18 November 2010 that includes 47 pages;
- Number R76/2006-NL1-10.43a dated 23 November 2010 that includes 18 pages;
- Number R76/2006-NL1-10.43b dated 23 November 2010 that includes 28 pages;
- Number NMI-11200439-01 dated 8 March 2012 that includes 34 pages;
- Number NMI-11200439-02 dated 8 March 2012 that includes 24 pages;
- Number NMI-11200439-03 dated 8 March 2012 that includes 29 pages;
- Number NMI-11200439-04 dated 8 March 2012 that includes 20 pages;
- Number NMI-11200439-05 dated 8 March 2012 that includes 19 pages;
- Number NMI-11200439-06 dated 8 March 2012 that includes 7 pages;
- Number NMI-11200439-07 dated 8 March 2012 that includes 25 pages;
- Number NMI-11200209-01 dated 8 March 2012 that includes 66 pages;
- Number NMI-11200209-02 dated 8 March 2012 that includes 49 pages;
- Number NMI-11200385-01 dated 5 January 2012 that includes 24 pages;
- Number NMI-11200385-02 dated 5 January 2012 that includes 21 pages;
- Number NMI-11200385-03 dated 5 January 2012 that includes 21 pages.

Characteristics of the non-automatic weighing instrument:

- Maximum number of scale intervals : $n \leq 100000$ for class $\textcircled{\text{II}}$ instruments or $n \leq 10000$ for class $\textcircled{\text{III}}$ instruments or $n \leq 1000$ for class $\textcircled{\text{III}}$ instruments.
- Maximum partial weighing ranges : 3
- Temperature range : $-10\text{ }^{\circ}\text{C} / +40\text{ }^{\circ}\text{C}$
- Tare : $T \leq \text{Max}$