

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-12.16 Revision 1  
Project number 12200014  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Shanghai Teraoka Electronic Co., Ltd. Ting Lin Industry Developmental Zone, Jinshan County, Shanghai 201505 P.R. of China
Manufacturer	Shanghai Teraoka Electronic Co., Ltd. Ting Lin Industry Developmental Zone, Jinshan County, Shanghai 201505 P.R. of China
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : DS-686
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
27 August 2012

C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-12.16 Revision 1  
Project number 12200014  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- Number R76/1992-NL1-10.12A that includes 56 pages;
- Number R76/1992-NL1-10.12B that includes 17 pages;
- Number R76/1992-NL1-10.44 that includes 14 pages;
- Number NMI-11200635-01 dated 30 September 2011 that includes 23 pages;
- Number NMI-12200014-01 dated 12 April 2012 that includes 30 pages.

### Characteristics of the non-automatic weighing instrument:

- Maximum capacity :  $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
- Verification scale interval :  $e \geq 1 \text{ g}$
- Maximum number of scale intervals :  $n \leq 6000$  divisions for single interval  
 $n \leq 3000$  divisions (per partial weighing range) for multi-interval
- Maximum partial weighing ranges : 2
- Temperature range :  $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
- Tare :  $T \leq -50\%$  for single interval instruments  
 $T \leq -\text{Max}_1$  for multi interval instruments

### Revision History

Revision	Date	Change(s)
Initial	12 April 2012	-
1	27 August 2012	Correction of the OIML R76 edition on page 1.