

OIML Certificate of Conformity

OIML Member State The Netherlands Number R76/2006-NL1-12.29 Project number 12200402 Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oos	• • • • • • • • • • • • • • • • • • •	
Applicant	Fook Tin Technologies Ltd 4F, Eastern Centre 1065 King's Rd. Quarry Bay Hong Kong		
Manufacturer	Fook Tin Technologies Ltd 4F, Eastern Centre 1065 King's Rd. Quarry Bay Hong Kong		
Identification of the certified type	A Non-automatic weigh Type	ing instrument : FMM-Fx, PD350x, PD150x,	
Characteristics	See next page	FMM-PDx, PD300x	
+ identified in the OIML		e identified Type (represented by the sample(s) rements of the following Recommendation of the ML):	
	OIML R76 - Edition 2006 f	for accuracy class III	
instrument covered by		d technical characteristics of the type of measuring onal Recommendation above-identified. international approval.	
OIML Member State in	which the Certificate was is	rtificate's reference number and the name of the sued, partial quotation of the Certificate and of d, although either may be reproduced in full.	
+ Issuing Authority	NMi Certin B.V., OIML Is 26 November 2012	suing Authority NL1 + + + + + + + + + + + + + + + + + + +	
	C. Oosterman		
	Head Certification Board		
NMi Certin B.V.	This document is issued under the	Parties concerned can	
Hugo de Grootplein 1	provision that no liability is	lodge objection against	
3314 EG Dordrecht the Netherlands	accepted and that the applicant shall indemnify third-party liability.	this decision, within six weeks after the date of	
T +31 78 6332332		submission, to the	
certin@nmi.nl www.nmi.nl	The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org	general manager of NMi (see www.nmi.nl).	



OIML Certificate of Conformity

OIML Member State The Netherlands Number R76/2006-NL1-12.29 Project number 12200402 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s): - No. NMi-12200402-01 dated 8 November 2012 that includes 54 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	
Maximum capacity	Max ≤ 250 kg
Verification scale interval	e ≥ 0,1 kg
Maximum number of scale intervals	n ≤ 2500 divisions
* * * * * * * * * * * *	+ + + (per partial weighing range) + + + +
Maximum partial weighing ranges	* * * * * * * * 1 * * * * * * * *
Temperature range	+10 °C / +40 °C
Tare	$T \leq$ -Max for single interval instruments
Weighing range(s)	Single interval
Power supply voltage	100 - 240 V AC, 50/60 Hz (with external adapter) 9 V DC (supplied by external adapter)
Software identification	+ + + + + Version number: U1.0 + + + + +
* * * * * * * * * * * * * * *	* * * * * * * * * * * * * * * * * * *