

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-12.42 revision 1 Project number 12200441 Page 1 of 2

NMi Certin B.V. Issuing authority

Person responsible: C. Oosterman

Mettler-Toledo LLC Applicant

1150 Dearborn Drive Worthington, Ohio 43085 United States of America

Mettler-Toledo LLC Manufacture

> 1150 Dearborn Drive Worthington, Ohio 43085 **United States of America**

Identification of the

certified type

A Non-automatic weighing inst

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class (III)



This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

NMi Certin B.V.

C. Oosterman Head Certification Board

NMi Certin B V Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-12.42 revision 1 Project number 12200441 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-03.16A dated 4 July 2003 that includes 53 pages;
- No. R76/1992-NL1-03.16B dated 4 July 2003 that includes 15 pages;
- No. 703861A dated 7 July 2007 that includes 12 pages;
- No. NMi-12200441-01 dated 10 January 2013 that includes 28 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	(III)
Maximum capacity	+ + + + + 3 kg ≤ Max ≤ 15 kg
Verification scale interval	+ + + + + + + e ≥ 1 g + + + + + + + + + + + + + + + + + +
Maximum number of scale intervals	n ≤ 3000 divisions (per partial weighing range)
Maximum partial weighing ranges	* * * * * * * * * 2 * * * * * * * * * *
Temperature range	0 °C / + 40 °C
Weighing range(s)	Single interval Multi-interval
Power supply voltage	100 – 240 V AC 50/60 Hz with external adapter 7,5 V DC to 15 V DC (supplied by an AC/DC adapter) 5 V DC from USB
Application	Intended to be used for direct sales to the public
Software identification	Version number: 23.19

Revision History

This revision replaces the previous version.

Revision	Date + + + +	Change(s)	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+ -
Initial+ +	27 February 2013	+ + + + + +	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+
1 + + +	11 March 2013	Correction of mar	nufa	ctu	rers	s na	me	e ar	nd a	рр	lica	ant	s na	ame	÷+	+	+	+ +