

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-13.13 Project number 12200371 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant Xiamen Pinnacle Electrical Co., Ltd.

4F, Guangxia Building, Torch High-Tech Zone, 361006

Xiamen, Fujian

China

Manufacturer Xiamen Pinnacle Electrical Co., Ltd.

4F, Guangxia Building, Torch High-Tech Zone, 361006

Xiamen, Fujian

China

Identification of the

certified type

A Non-automatic weighing instrument

Type

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class (III)



Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

issuing Authority Nivil Certin B.V., Olivic issuing Authority NL i

This Certificate does not bestow any form of legal international approval.

23 May 2013

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member StateThe Netherlands

Number R76/2006-NL1-13.13 Project number 12200371 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMi-10200906-01 dated 20 April 2011 that includes 34 pages;
- No. NMi-12200371-01 dated 21 May 2013 that includes 35 pages.

Characteristics of the non-automatic weighing instrument:

| Accuracy class | (II) |
|---|---|
| Maximum capacity | 6 kg ≤ Max ≤ 30 kg |
| Verification scale interval + + + + + + | $+ + + + + + + + e \ge 2g + + + + + + + + + + + + + + + + + + $ |
| Maximum number of scale intervals | n ≤ 3000 divisions (per partial weighing range) |
| Maximum partial weighing ranges | + + + + + + + + 2 + + + + + + + + |
| Temperature range | 0 °C / +40 °C |
| Tare | $T \le -(Max_1 - e_1)$ for multi interval instruments |
| Weighing range(s) | Multi-interval |
| Power supply voltage | 100 – 240 V AC 50/60 Hz 5 V DC |
| Application + + + + + + + + + + | Intended to be used for direct sales to the public |

5