

OIML Member State
The Netherlands

Number R76/2006-NL1-13.14
Project number 13200104
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Mettler-Toledo, LLC 1150 Dearborn Drive Worthington, Ohio 43085 United States of America
Identification of the certified type	A Non-automatic weighing instrument Type : PS15
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
9 December 2013



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



OIML Member State
The Netherlands

Number R76/2006-NL1-13.14
Project number 13200104
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-03.16A that includes 53 pages;
- No. R76/1992-NL1-03.16B that includes 15 pages;
- No. NMI-13200104-01, dated 3 December 2013 that contains 28 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Maximum number of scale intervals	$n \leq 3000$ divisions (per partial weighing range)
Maximum partial weighing ranges	2
Temperature range	0 °C / + 40 °C
Weighing range(s)	Single interval Multi-interval
Power supply voltage	100 V – 240 V AC 50/60 Hz with external adapter 7,5 V – 15 V DC (supplied by an AC/DC adapter)
Application	Intended to be used for direct sales to the public
Software identification	Version number: 2.29