

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-13.41 Project number 13200425 Page 1 of 2

NMi Certin B.V. Issuing authority

Person responsible: C. Oostermar

Applicant and Manufacturer

Mettler-Toledo (Changzhou) Measurement Technology Ltd. No. 111, West Taihu Road, ChangZhou XinBei District

213125 Jiangsu

Peoples Republic of China

Identification of the certified type

A Non-automatic weighing instrument

CUB II (RWXX... XRWXX... BPA224)

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76-1, Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

NMi Certin B.V., OIML Issuing Authority

29 October 2013

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-13.41 Project number 13200425 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-03.49A dated 6 November 2003 that includes 56 pages;
- No. R76/1992-NL1-03.49B dated 6 November 2003 that includes 44 pages;
- No. R76/1992-NL1-07.13 dated 7 May 2007 that includes 36 pages;
- No. NMi-13200425-01 dated 25 October 2013 that includes 58 pages;
- No. NMi-13200425-02 dated 22 October 2013 that includes 28 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class + + + + + + + + + +	+ + + + + + + + + + + + + + + + + + + +
Maximum capacity	1,5 kg ≤ Max ≤ 15 kg
Verification scale interval	e ≥ 0,5 g
Maximum number of scale intervals	n ≤ 3000 divisions
Temperature range + + + + + + +	+ + + + + + -10 °C / +40 °C + + + + + + +
Tare + + + + + + + + + + + + +	T ≤ -Max
Weighing range(s)	Single interval
Power supply voltage	9 V DC through AC/DC adapter 6 x 1,5 V batteries 6 V DC rechargeable battery
Application	Intended to be used for direct sales to the public
Software identification	Version number: L3.1 or any previous version number

Software:

- The identification number will be displayed at start-up;
- The indicator has embedded software.