

OIML Member State
The Netherlands

Number R76/2006-NL1-14.08
Project number 13200752
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Ohaus Corporation 7 Campus Drive, Suite 310, Parsippany, NJ 07054 United States of America
Identification of the certified type	A Non-automatic weighing instrument Type : Explorer EX...series
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **I** or **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
25 March 2014

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-14.08
Project number 13200752
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- Number NMI-11200007-01 dated 14 April 2011 that includes 40 pages;
- Number NMI-11200007-02 dated 14 April 2011 that includes 28 pages;
- Number NMI-11200007-03 dated 14 April 2011 that includes 20 pages;
- Number NMI-11200007-04 dated 14 April 2011 that includes 31 pages;
- Number NMI-11200007-05 dated 14 April 2011 that includes 20 pages;
- Number NMI-13200207-01 dated 16 July 2013 that includes 35 pages;
- Number NMI-13200752-01 dated 28 February 2014 that includes 21 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	I or II
Maximum capacity	220 g ≤ Max ≤ 35000 g or 1100 ct ≤ Max ≤ 175000 ct
Verification scale interval	1 mg ≤ e < 1 g or 0,005 ct ≤ e < 5 ct
Maximum number of scale intervals	n ≤ 420000 divisions
Temperature range	+10 °C / +30 °C
Tare	T ≤ -Max
Weighing range(s)	Single interval
Power supply voltage	100 – 240 V AC, 50/60 Hz or AC/DC Adapter: input 100 - 240 V AC, 50/60 Hz, output 12 V DC
Software identification	Base: 1.xx or 2.xx Terminal: 1.xx or 2.xx Where xx can be 00-99, represents the non legally relevant parts. 1. or 2. represent the legal-relevant part of the software