

OIML Member State
The Netherlands

Number R76/2006-NL1-14.25
revision1
Project number 14200044
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Mettler-Toledo Instrument (Shanghai) Co., Ltd. 589 GuiPing Road Shanghai 200233 Peoples Republic of China
Identification of the certified type	A Non-automatic weighing instrument Type : HE... Moisture Analyzer
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
17 July 2014



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-14.25
revision1
Project number 14200044
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-14200044-01 revision 1 dated 16 July 2014 that includes 52 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	II
Maximum capacity	$54 \text{ g} \leq \text{Max} \leq 81 \text{ g}$
Verification scale interval	$e \geq 10 \text{ mg}$
Maximum number of scale intervals	$n \leq 8100 \text{ divisions}$
Temperature range	$+10 \text{ }^\circ\text{C} / +30 \text{ }^\circ\text{C}$
Tare	$T \leq -\text{Max}$
Weighing range(s)	Single interval
Power supply voltage	100 - 130 V AC 50/60 Hz 200 - 240 V AC 50/60 Hz
Software identification	Version number: 1.xx, where xx is a number between 10 and 99 that represents metrological not relevant changes

Revision History

This revision replaces the previous version(s).

Revision	Date	Change(s)
Initial	8 July 2014	-
1	17 July 2014	Name change to include "Moisture Analyzer".