

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-14.58 Project number 14200417 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Ooster

Applicant and Manufacturer

Mettler-Toledo, LLC

1150 Dearborn Drive

Worthington, Ohio 43085-671

United States of America

Identification of the certified type

A Non-automatic weighing instrument

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

NMi Certin B.V., OIML Issuing Authority

11 November 2014

NMi Certin B V Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-14.58 Project number 14200417 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMi-14200417-01 dated 16 October 2014 that includes 29 pages;
- No. NMi-14200417-02 dated 16 October 2014 that includes 29 pages.

Characteristics of the non-automatic weighing instrument:

Non-automatic weighing instrument		MT1260	0795	SLP331D
+ + + + + + + + + +		+ + + + + + +		+ + + + + + +
Accuracy class				
Maximum capacity		Max ≤ 150 kg	Max ≤ 150 kg	30 ≤ Max ≤ 50 kg
Verification scale interval		e ≥ 0,05 kg	e ≥ 0,02 kg	e ≥ 0,001 kg
Weighing range(s)		Single interval	Single interval Multiple range	Single interval Multi-interval Multiple range
Maximum number of scale intervals (one weighing range)		n ≤ 3000 divisions	n ≤ 3000 divisions	n ≤ 3500 divisions
Maximum number of scale + + intervals (multi-interval) + + +			· · · · · · · · ·	n ≤ 3500 divisions (per partial weighing range)
Maximum number of partial weighing ranges		+ + + + + + + +	+ + + + + + + + + + + + + + + + + + +	+ + + 3 + + +
Maximum number of scale intervals (multiple range)			n ≤ 3000 divisions (per weighing range)	n ≤ 3500 divisions (per weighing range)
Maximum numb	per of weighing	+ + + +1+ + + +	+ + + +2+ + + +	+ + + -3 + + +
Tare		T ≤ -Max		
Temperature range		0 °C / 40 °C	0 °C / 40 °C	0 °C / 40 °C
Power supply voltage		5 V DC (USB)		
Application •	+ + + + +		e used for determining a tra b be used for direct sales to t	
Software	Version	0.00.xxxx Where xxxx is a number between 0000 and 9999 which represents the non legally relevant part of the software		
identification	Checksum (CRC32)	0x93B46AB4	0x805D6909	0x9187A86A