

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-14.61 Project number 13200324 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Ooster

Applicant and

Ohaus Corporation

Manufacturer

7 Campus Drive, Suite 310 Parsippany, NJ 07054 United States of America

Identification of the certified type

A Non-automatic weighing instrument

S71 series

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

NMi Certin B.V., OIML Issuing Authority

10 November 2014

NMi Certin B V Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State The Netherlands

Number R76/2006-NL1-14.61 Project number 13200324 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMi-13200324-01 dated 29 November 2013 that includes 59 pages;
- No. NMi-13200324-02 dated 15 October 2014 that includes 14 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	+ + + + + + + + + + + + + + + + + + + +
Maximum capacity	3 kg ≤ Max ≤ 30 kg
Verification scale interval	+ + + + + + + e ≥ 1 g + + + + + + + + + + + + + + + + + +
Weighing range(s) + + + + + + + +	+ + + + + + + Single interval + + + + + +
	+ + + + + Multi-interval + + + + + +
Maximum number of scale intervals (multi-interval)	$n \le 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	+ + + + + + + + + 2 + + + + + + + + + +
Temperature range	+ + + + + + -10 °C / +40 °C + + + + + + +
Tare + + + + + + + + + + + + + + + + + + +	T ≤ -Max
Power supply voltage + + + + + + +	100 – 240 V AC 50/60 Hz 12 V rechargeable battery
Application	Intended to be used for direct sales to the public
Software identification scale application	Version: "1.xxx.yyy" where "xxx" represents the legally non-relevant software part and "yyy" represents bug fixes.

5