

OIML Certificate of Conformity

OIML Member State The Netherlands Number R76/2006-NL1-15.05 Project number 14200704 Page 1 of 2

NMi Certin B.V. Issuing authority Person responsible: C. Oost InBody Co., Ltd. Applicant and Manufacturer InBody Bldg., 54, Nonhyeon-ro 2-gil Gangnam-gu, Seoul 135 Korea Identification of the A Non-automatic weighing instrument certified type Type InBody J30 / 230 / 270/ Characteristics See next page This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML): OIML R 76 - Edition 2006 for accuracy class This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval. Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full. NMi Certin B.V., OIML Issuing Authority Issuina 10 February 2015 Sosterman Head Certification Board NMi Certin B V This document is issued under the Parties concerned can Hugo de Grootplein 1 provision that no liability is lodge objection against 3314 EG Dordrecht accepted and that the applicant this decision, within six the Netherlands shall indemnify third-party liability. weeks after the date of T+31 78 6332332 submission, to the The notification of NMi Certin B.V. general manager of NMi certin@nmi.nl as Issuing Authority can be verified www.nmi.nl (see www.nmi.nl). at www.oiml.org



OIML Certificate of Conformity

OIML Member State The Netherlands Number R76/2006-NL1-15.05 Project number 14200704 Page 2 of 2

	March 2012 that includes 30 pages; Ianuary 2015 that includes 18 pages. natic weighing instrument:
Accuracy class	
Maximum capacity	Max ≤ 200 kg
Verification scale interval	+ + + + + + + + + + + + + + + + + + +
Weighing range	Single interval
Maximum number of scale intervals	$n \le 2000$ divisions
Temperature range	+ + + + + + + 0 °C / +40 °C
Power supply voltage + + + + +	12 V DC through 100 - 240 V AC 50/60 Hz adapter
Software identification	Version numbers: IB230-1.0 / IB270-1.0 / IB470-1.0 / IBJ30-1.0