

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-15.12 Revision 1  
Project number SO15202444  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Shinko Denshi Co., Ltd. 3-9-11 Yushima, Bunkyo-ku, Tokyo, 113-0034 Japan
Identification of the certified type	A <b>Non-automatic weighing instrument</b> Type : AB, RJ
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76-1**, Edition 2006 for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
23 June 2015



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-15.12 Revision 1  
Project number SO15202444  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-13200783-01 dated 24 April 2014 that includes 43 pages;
- No. NMI-13200783-02 dated 24 April 2014 that includes 20 pages;
- No. NMI-13200783-03 dated 24 April 2014 that includes 30 pages;
- No. NMI-13200783-04 dated 12 March 2014 that includes 17 pages;
- No. NMI-15200108-01 dated 2 April 2015 that includes 12 pages;
- No. NMI-15200108-02 dated 2 April 2015 that includes 16 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	II
Maximum capacity	$320 \text{ g} \leq \text{Max} \leq 12000 \text{ g}$
Verification scale interval	$e \geq 0,01 \text{ g}$
Actual scale interval	$e = d \text{ or } e = 10 d$
Weighing range	Single interval
Maximum number of scale intervals	$n \leq 62000 \text{ divisions}$
Tare	$T \leq -\text{Max}$
Temperature range	+5 °C / +40 °C
Power supply voltage	6 V DC through 100 – 240 V 50/60 Hz AC/DC adapter or 4x 1,5 V AA size battery
Software identification	Checksum: EE52 or 0689 or D31C

## Revision History

This revision replaces the previous version.

Revision	Date	Changes
0	16 April 2015	Initial
1	23 June 2015	New software identification added, due to minor change in software.