

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-15.26  
Project number 15200070  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Dräger Medical Systems, Inc. 3135 Quarry Road Telford, PA 18969 United States of America
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : Resuscitaire Radiant Warmer (RW82, RW82-VHA) Resuscitaire Birthing Room Warmer (WBR82)
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
10 September 2015



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report:  
- No. NMI-15200070-01 dated 10 September 2015 that includes 56 pages.

**Characteristics of the non-automatic weighing instrument:**

Accuracy class	III
Maximum capacity	Max $\leq$ 8 kg
Verification scale interval	e $\geq$ 10 g
Weighing range	Single interval
Maximum number of scale intervals	n $\leq$ 800 divisions
Tare	T $\leq$ -4 kg
Temperature range	+10 °C / +40 °C
Power supply voltage	12 V DC through an AC/DC power supply of 100 – 240 V AC 50/60 Hz
Application	To be used only for weighing patients
Software identification	Version number: 1.08