

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-15.54 Project number SO15204385 Page 1 of 2

NMi Certin B.V. Issuing authority

Person responsible: C. Oosterma

Applicant and Manufacturer

Shinko Denshi Co., Ltd. 3-9-11 Yushima, Bunkyo-ku,

Tokyo, 113-0034

Japan

Identification of the certified type

A Non-automatic weighing instrument

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76-1, Edition 2006 for accuracy class (II)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

NMi Certin B.V., OIML Issuing Authority

10 December 2015

NMi Certin B V Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-15.54 Project number SO15204385 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMi-13200783-01 dated 24 April 2014 that includes 43 pages;
- No. NMi-13200783-02 dated 24 April 2014 that includes 20 pages;
- No. NMi-13200783-03 dated 24 April 2014 that includes 30 pages;
- No. NMi-13200783-04 dated 12 March 2014 that includes 17 pages;
- No. NMi-15200108-01 dated 2 April 2015 that includes 12 pages;
- No. NMi-15200108-02 dated 2 April 2015 that includes 16 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	+ + + + + + + + + + + + + + + + + + + +
Maximum capacity + + + + + + +	+ + + + + 320 g ≤ Max ≤ 12000 g + + + + +
Verification scale interval	+ + + + + + + + e ≥ 0,01 g + + + + + + + + + + + + + + + + + +
Actual scale interval	e = d or e = 10 d
Weighing range	Single interval
Maximum number of scale intervals	n ≤ 62000 divisions
Tare + + + + + + + + + + + +	+ + + + + + + + T ≤ -Max+ + + + + + + + + + + + + + + + + + +
Temperature range	+5 °C / +40 °C
Power supply voltage + + + + + -	6 V DC through 100 – 240 V 50/60 Hz AC/DC adapter or 4x 1,5 V AA size battery
Software identification	Checksum: EE52, 0689, D31C, 803A

5