

OIML Member State
The Netherlands

Number R76/2006-NL1-16.01
Project number 14200507
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Nagata Scale Co.Ltd. No. 3, Lane 404, Chung Cheng S. Rd. Yung Kang Dist. Tainan City Taiwan
Identification of the certified type	A Non-automatic weighing instrument Type : BW-0365
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
29 January 2016

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-14200507-01 dated 29 January 2016 that includes 16 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 20 \text{ kg}$
Verification scale interval	$2 \text{ g} \leq e \leq 10 \text{ g}$
Weighing ranges	Single interval Multi-interval Multiple range
Maximum number of scale intervals (one weighing range)	$n \leq 3000$ divisions
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Maximum number of scale intervals (multiple range)	$n \leq 3000$ divisions (per weighing range)
Maximum number of weighing ranges	2
Tare	$T \leq -\text{Max}$
Temperature range	+5 °C / +40 °C
Power supply voltage	6 V DC through an AC/DC plug-in power supply of 230 V AC 50 Hz, or 4,5 V DC (by 2 sets of AA size internal batteries)
Application	To be used only for weighing patients
Software identification	Version number: 83-xx (xx is a number between 00-99 and represents the non-legally relevant software)