

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-16.04 Project number 14200409 Page 1 of 2

NMi Certin B.V. Issuing authority

Person responsible: C. Oostermar

Applicant and

Rice Lake Weighing Systems

Manufacturer

230 W Coleman St. Rice Lake, WI 54868

United States of America

Identification of the

An Indicator

See next page

certified type

Type 1280 Enterprise Series

Characteristics

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III) (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

NMi Certin B.V., OIML Issuing Authority

29 February 2016

NMi Certin B V Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-16.04 Project number 14200409 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMi-14200409-01 dated 29 February 2016 that includes 22 pages;
- No. NMi-14200409-02 dated 29 February 2016 that includes 50 pages;
- No. NMi-14200409-03 dated 29 February 2016 that includes 22 pages.

Characteristics of the indicator:

Accuracy class	Or (III)
Weighing ranges	Single interval Multi-interval Multiple range
Maximum number of scale intervals (one weighing range)	+ + + + + n ≤ 10000 divisions + + + + +
Maximum number of scale intervals (multi-interval)	n ≤ 10000 divisions (per partial weighing range)
Maximum number of partial weighing ranges	+ + + + + + + + 3+ + + + + + + + +
Maximum number of scale intervals (multiple range)	n ≤ 10000 divisions (per weighing range)
Maximum number of weighing ranges	+ + + + + + + + + + + + + + + + + + + +
Load cell excitation voltage	+ + + + + + 10 V DC + + + + + +
Minimum input voltage per verification scale interval + + + + + + + + + + + + + + + + + + +	+ + + + + + + + + + + + + + + + + + +
Minimum load cell resistance	+ + + + + + + 23 Ω + + + + + + + +
Maximum load cell resistance	1050 Ω
Fraction of the maximum permissible error	0,5
Load cell connection	6-wire (remote sensing)
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells	395 m/mm ² In case a 4-wire connection is used, the load cells are connected directly without junction box
Maximum number of load platforms + + + +	+ + + + + + + + + + + + + + + + + + + +
Temperature range	+ + + + + + + -10 °C / +40 °C
Power supply voltage	100 – 240 V AC 50/60 Hz (for AC version) 9 – 30 V DC (for DC version)
Software identification + + + + + + + + + + + + + + + + + + +	Version number: V1.xx (xx is a number between 00 and 99 and represents the non-legally relevant software)