



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-16.25
Project number SO16201036
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Ohaus Corporation 7 Campus Drive, Suite 310 Parsippany, NJ 07054 United States of America
Identification of the certified type	A Non-automatic weighing instrument Type : RS-series
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
29 August 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-NL1-16.25
Project number SO16201036
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-05.16 dated 3 June 2005 that includes 55 pages;
- No. R76/1992-NL1-06.38 dated 2 November 2006 that includes 44 pages;
- No. R76/1992-NL1-09.12 dated 6 April 2009 that includes 17 pages;
- No. NMI-12200063-01 dated 25 May 2012 that includes 26 pages.

Characteristics of the non-automatic weighing instrument:

Load cell	AMI	New AMI
Accuracy class	III	
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval	
Maximum number of scale intervals	$n \leq 3000$ divisions	$n \leq 6000$ divisions
Tare	$T \leq -\text{Max}$	$T \leq -\text{Max}$
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$	
Power supply voltage	100 - 240 V AC 50/60 Hz 12 V optional internal rechargeable battery	
Application	Intended to be used for direct sales to the public	
Software identification	179403; A 1.xx (xx = 00 ...99)	

The software identification is displayed at start-up;
The software is embedded.