

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-16.30 Project number 14200678 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and Manufacturer

Mettler-Toledo (Albstadt) GmbH

Unter dem Malesfelsen 34 72458 Albstadt

Germany

Identification of the certified type

A Non-automatic weighing instrument

Type : FreshWay :

FW .

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

12 April 2016

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-16.30 Project number 14200678 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMi-14200678-01 dated 14 July 2015 that includes 50 pages;
- No. NMi-14200678-02 dated 13 July 2015 that includes 8 pages;
- No. NMi-14200678-03 dated 8 April 2016 that includes 9 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		+++++++++++++++++++++++++++++++++++++++
Maximum capacity		6 kg ≤ Max ≤ 30 kg
Verification scale interval		e ≥ 2 g for single interval e ≥ 1 g for multi-interval
Weighing range(s)		Single interval Multi-interval
Maximum number of scale intervals (one weighing range)		n ≤ 6000 divisions
Maximum number of scale intervals (multi-interval)		n ≤ 3000 divisions (per partial weighing range)
Maximum number of partial weighing + + ranges + + + + + + + + + + + + + + + + + + +		+ + + + + + + + + 2 + + + + + + + + + +
Tare + + + + + + + + + + + + +		+ + + + + + + T < -Max + + + + + + + +
Temperature range		+ + + + + + -10 °C / +40 °C
Power supply voltage		100 – 240 V AC 50/60 Hz
Application		Intended to be used for direct sales to the public
Software didentification	+ + + + + Name+	MT UPOSScale
	+ + + Version number	1.x.y (x = 0999, y = 0999)

5