

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-16.37 Project number SO16201323 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Ooster

Applicant and

Ohaus Corporation

Manufacturer

7 Campus Drive, Suite 310

Parsippany, NJ 07054 United States of America

Identification of the certified type

A Non-automatic weighing instrument

Characteristics

See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)



This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

NMi Certin B.V., OIML Issuing Authority

19 May 2016

NMi Certin B V Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-16.37 Project number SO16201323 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/1992-NL1-06.18A dated 25 April 2006 that includes 44 pages;
- No. R76/1992-NL1-06.18B dated 25 April 2006 that includes 28 pages;
- No. R76/1992-NL1-07.11A dated 21 March 2007 that includes 16 pages;
- No. R76/1992-NL1-07.11B dated 19 July 2007 that includes 26 pages;
- No. R76/1992-NL1-09.30 dated 3 October 2009 that includes 12 pages;
- No. 811744 dated 23 February 2007 that includes 12 pages;
- No. NMi-12200063-03 dated 25 May 2012 that includes 26 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		
	6 kg ≤ Max ≤ 30 kg	6 kg ≤ Max ≤ 1500 kg
Maximum capacity	+ + + + + + + + +	(When the instrument is used with scale bases
		using the IDNet scale interface)
Verification scale interval	e ≥ 1 g	
Weighing range(s)	Single interval Multi-interval	
Maximum number of scale intervals (one weighing range)	n \leq 6000 divisions (e \geq 2 g)	
Maximum number of scale intervals (multi-interval)	$n \le 3000$ divisions (e ≥ 1 g) (per partial weighing range)	
Maximum number of partial weighing ranges	+ + + + + + + + 2	
Tare + + + + + + + + + + + + + +	+ + + + + + + + T ≤ -Max + + + + + + + + + + + + + + + + + + +	
Temperature range	+ + + + + + -10 °C/+40 °C + + + + + +	
Power supply voltage	100 – 240 V AC 50/60 Hz	
Application	Intended to be used for direct sales to the public	

5