

## OIML Certificate of Conformity

**OIML Member State** 

The Netherlands

Number R76/2006-NL1-16.44 Project number 16200406 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and Manufacturer

Teraoka Seiko Co., Ltd. 13-12 Kugahara 5-Chome Ohta-Ku, Tokyo 146-8580

Japan

Identification of the certified type

A Non-automatic weighing instrument

Type : DPS-560

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

19 July 2016

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







## OIML Certificate of Conformity

**OIML Member State** 

The Netherlands

Number R76/2006-NL1-16.44 Project number 16200406 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMi-12200108-01 dated 25 June 2014 that includes 49 pages;
- No. NMi-12200108-04 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-06 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-13 dated 13 May 2013 that includes 15 pages;
- No. NMi-12200108-14 dated 14 May 2014 that includes 10 pages;
- No. NMi-14200477-01 dated 26 September 2014 that includes 15 pages;
- No. R76/1992-NL1-10.11 revision 1 dated 9 January 2015 that includes 26 pages;
- No. NMi-15200354-01 dated 31 August 2015 that includes 17 pages;
- No. NMi-15200743-01 dated 13 May 2016 that includes 8 pages;
- No. NMi-16200406-01 dated 15 July 2016 that includes 12 pages.

## Characteristics of the non-automatic weighing instrument:

++++++
6 kg ≤ Max ≤ 30 kg
e ≥ 1 g
Single interval Multi-interval
† † † † † n ≤ 3000 divisions
n ≤ 3000 divisions (per partial weighing range)
+ + + + + + + + + + + + + + + + + + + +
$T \le -Max$ for instruments with single interval $T \le -Max_1$ for multi-interval instruments
+ + + + + + -10 °C / +40 °C + + + + + + +
100 - 240 V AC 50/60 Hz
Intended to be used for the making-up of prepackages
See certificate TC8109
+ + + + See certificate TC8591 + + + + +

5