

OIML Member State
The Netherlands

Number R76/2006-NL1-16.46
Project number 15200291
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Xiamen Jadever Scale Co., Ltd. No. 40 Huli Avenue, Huli District Xiamen 361006, Fujian China
Identification of the certified type	A non-automatic weighing instrument Type : JWL
Characteristics	See next page


This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
19 August 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



OIML Member State
The Netherlands

Number R76/2006-NL1-16.46
Project number 15200291
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Report:

- No. NMI-15200291-01 dated 25 July 2016 that includes 47 pages.

Characteristics of the indicator type JWI-700W:

Accuracy class	Ⓜ or ⓂⓂ
Weighing range	Single interval
Maximum number of scale intervals	$n \leq 3000$ divisions
Load cell excitation voltage	5 V DC
Minimum input voltage per verification scale interval	2,0 μ V
Minimum load cell resistance	40 Ω
Maximum load cell resistance	1215 Ω
Fraction of the maximum permissible error	0,5
Load cell connection	4-wire
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells	The load cells are connected directly without junction box
Temperature range	-10 °C / +40 °C
Power supply voltage	230 V AC 50 Hz, or 6 V DC battery
Software identification	Version number: v1.19

The non-automatic weighing instrument is built from the indicator with load cell and construction according to clause 3.10.2 and annex F.