OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.05
Project number 16200470

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Teraoka Seiko Co., Ltd. 5-13-12, Kugahara, Ohta-ku, 146-8580 Tokyo Japan
Identification of the certified type	A **Non-automatic weighing instrument** Type : DPS-5600, DPS-5600M
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
23 March 2017

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.05
Project number 16200470

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMi-12200108-03 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-04 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-05 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-06 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-13 dated 13 May 2013 that includes 15 pages;
- No. NMi-12200108-01 dated 25 June 2014 that includes 49 pages;
- No. R76/1992-NL1-10.11 revision 1 dated 9 January 2015 that includes 26 pages;
- No. NMi-SO15202422-01 dated 19 August 2015 that includes 9 pages;
- No. NMi-15200354-01 dated 31 August 2015 that includes 17 pages;
- No. NMi-16200470-01 dated 22 March 2017 that includes 27 pages;
- No. NMi-16200470-02 dated 22 March 2017 that includes 27 pages;
- No. NMi-16200470-04 dated 22 March 2017 that includes 47 pages;
- No. NMi-16200470-05 dated 22 March 2017 that includes 12 pages;
- No. NMi-16200470-06 dated 22 March 2017 that includes 17 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		(III)
Maximum capacity		3 kg ≤ Max ≤ 150 kg
Verification scale interval		e ≥ 1 g
Weighing ranges		Single interval Multi-interval
Maximum number of scale intervals (one weighing range)		n ≤ 3000 divisions
Maximum number of scale intervals (multi-interval)		n ≤ 3000 divisions (per partial weighing range)
Maximum number of partial weighing ranges		2
Tare		T ≤ -Max for instruments with one weighing range T ≤ -Max_1 for multi-interval instruments
Temperature range		-10 °C / +40 °C
Power supply voltage		100 – 240 V AC 50/60 Hz
Application		Intended to be used for the making-up of prepackages
Software identification	Measurement software	Version number: 1.xx or 2.xx or 3.xx (xx = 00 ... 99)
	A/D board software	Version numbers: 3.xx (For TPB-3356 A/D-board, xx = 22 ... 99), or 1.xx (For TPB-3772 A/D-board, xx = 00 ... 99)