

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-17.08 Revision 1 Project number 16200755 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and

Radwag Wagi Elektroniczne

Manufacturer

Bracka 28 26-600 Radom

Poland

Identification of the

A Non-automatic weighing instrument

certified type

Type + + + + + + + + + + : AS...R2.../AS...X2...

PS...R2... / PS...X2...

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class or II

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

25 April 2017

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/2006-NL1-17.08 Revision 1 Project number 16200755 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMi-16200755-01 dated 13 March 2017 that includes 29 pages;
- No. NMi-16200755-02 dated 13 March 2017 that includes 50 pages.

Characteristics of the non-automatic weighing instrument:

	<u> </u>	<u> </u>
	ASR2 / ASX2	PSR2 / PSX2
Accuracy class	1 + + + + + + + + + + + + + + + + + + +	
Maximum capacity	Max ≤ 310 g, or Max ≤ 1550 ct	Max ≤ 6000 g, or Max ≤ 30000 ct
Verification scale interval	$e \ge 1$ mg, or $e \ge 0.01$ ct	e ≥ 10 mg, or e ≥ 0,1 ct
Actual scale interval	e = d, or e = 10 d, or e = 100 d	
Weighing range(s)	Single interval	Single interval, or Multiple range
Maximum number of scale intervals (one weighing range)	n ≤ 310000 divisions	n ≤ 100000 divisions
Maximum number of scale intervals (multiple range)	+ + + + + + + + + + + + + + + + + + + +	n ≤ 20000 divisions (per weighing range)
Maximum number of weighing ranges	+ + + + + + + + + +	+ + + + + + + + + + + + + + + + + + + +
Tare+ + + + + + + + + + + + + + + + + + +	+ + + + + + + T ≤ -Max + + + + + + + + + + + + + + + + + + +	
Temperature range	+ + + + + + +10°C/+40°C + + + + + +	
Power supply voltage	100 – 240 V AC 50/60 Hz	
Software identification	Version numbers for weighing firmware: 2.2.1 (for X2 models), 4.0.1 (for R2 models)	

Revision History

This revision replaces the previous version.

Revision	Date + + + +	Changes + + + + + + + + + + + + + + + + + + +	
Initial	13 March 2017		
† + + + † + + +	25 April 2017	Reference to test report NMi-16200755-04 is removed.	