

OIML Certificate N° R85/1998-NL1-07.03

Project number 511252 Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name:

NMi Certin B.V.

Address:

Hugo de Grootplein 1, 3314 EG Dordrecht The Netherlands

Person responsible:

Ing. C. Oosterman

Applicant

Name:

Enraf B.V.

Address:

Delftechpark 39 2628 XJ Delft

The Netherlands

Manufacturer of the certified type

Name:

Enraf B.V.

Address:

Delftechpark 39 2628 XJ Delft

The Netherlands

Identification of certified type

Automatic level gauge for measuring the level of liquid in storage tanks, model Smartradar Flexline, type XP,

with the antennas F08, W06, H04, S06, S08, S10 and S12.

Antenna type	Accuracy class	Minimum product level	Maximum product level	Minimum and maximum values for liquid pressure, for liquid temperature and for liquid properties.	Minimum and maximum values for vapour pressure, for vapour temperature and for vapour properties.
F08, W06	2	20 m below the antenna	1 m below the antenna	The manufacturer shall specify these values for each application.	The manufacturer shall specify these values for each application.
H04, S06, S08, S10, S12	2	21 m below the antenna	1 m below the antenna	The manufacturer shall specify these values for each application.	The manufacturer shall specify these values for each application.



OIML Certificate N° R85/1998-NL1-07.03

Project number 511252 Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R85

The Netherlands

Edition 1998 (E) for accuracy class 2

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated **Test Report:**

R85/1998-NL1-07.03 that includes 100 pages.

The Issuing Authority Ing. C. Oosterman

Manager Product Certification

1 October 2007

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.