# **Physikalisch-Technische Bundesanstalt**

### Braunschweig und Berlin

Member State of OIML Germany



OIML Certificate N° R107/1997-DE1-05.01

### OIML CERTIFICATE OF CONFORMITY

#### **Issuing Authority**

Name:	Physikalisch-Technische Bundesanstalt
Address:	Bundesallee 100, 38116 Braunschweig
Person responsible:	Dr. Roman Schwartz

#### Applicant

Name:	CHROTEC GmbH
Address:	Josef - Dietzgen - Strasse 12, 53773 Hennef GERMANY

Manufacturer of the certified type is the applicant.

Identification of the	Discontinuous Totalizing Automatic Weighing Instrument
certified type	Type: Bulk Controller CHROTEC
	Accuracy class $\ge 0.2$ n $\le 6000$ (n $\le 3000$ per partial weighing range with multi-interval instruments) Max 3 kg to 300 t

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R107-1**, edition 1997 for accuracy class 0.2; 0.5; 1 and 2 (2 only for special purposes)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

## Physikalisch-Technische Bundesanstalt

OIML Certificate N° **R107/1997-DE1-05.01** 

The conformity was established by the results of tests and examinations provided in the associated Report 1.12-4017739 (15 pages) and Test Report No. 1.12-4017739/1 (50 pages).

#### The Issuing Authority

#### The CIML Member

Dr. R. Schwartz Direktor und Professor Prof. Dr. M. Kochsiek Vizepräsident

15.09.2005

15.09.2005

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Report is not permitted, although either may be reproduced in full.