Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML Germany



OIML Certificate N° R107/2007-DE1-09.01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt Address: Bundesallee 100, 38116 Braunschweig

Person responsible: Dr. Panagiotis Zervos

Applicant

Name: Bizerba GmbH & Co. KG

Address: Wilhelm-Kraut-Str. 65, 72336 Balingen

GERMANY

Manufacturer of the certified type is the applicant.

Identification of the cer-

tified type

Automatic discontinuous totalizing weighing instrument

Type: ST-Bulk

Accuracy class ≥ 0.2

 $n \le 6000$ ($n \le 3000$ per partial weighing range with multi-interval instru-

ments)

Max 3 kg to 300 t

Further characteristics see Report 1.12-4041913

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R107-1, edition 2007

for accuracy class 0.2, 0.5, 1 or 2

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not be tow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N° R107/2007-DE1-09.01

The conformity was established by the results of tests and examinations provided in the associated Report No. 1.12-4041913 that includes 16 pages and Test Report No. 1.12-4041913/1 that includes 57 pages.

The Issuing Authority

The CIML Member

Dr. P. Zervos Direktor und Professor Dr. R. Schwartz Direktor und Professor

18.11.2009

18.11.2009

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.