



**FEDERAL AGENCY FOR TECHNICAL REGULATION
AND METROLOGY OF RUSSIA**

OIML Member State
The Russian Federation

OIML Certificate №
R 117/1995-RU1-04.01
Revision 1-01.09.2009
The revision replaces the preceding one

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: VNIIMS
Address: 46, Ozernaya str, G-361, Moscow, 119361,
the Russian Federation

Person responsible: Dr. S.A. Kononogov

Applicant

Name: Larsen & Toubro Limited (L&T)
Petrol Dispensing Pumps & Systems
Address: Gate No.1, "C" Building 1st Floor
Powai works, Mumbai 400072, India

Manufacturer of the certified pattern is the Applicant

Identification of the certified pattern: L&T Fuel Dispensing Pump SPRINT series /Z-line series
/ PaceMaker-IV series

Further characteristics are set out on page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the sample identified in the associated test reports) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 117
edition 1995
for accuracy class 0.5

R 118
edition 1995

The revision replaces the preceding one

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in associated test report number 401-04/01 which includes pages.

The issuing authority

VNIIMS
Dr. S.A. Kenonogov
Director
Date: 01.09.2009



The CIML member

Dr. L. K. Issaev
Professor
Date: 01.09.2009

Characteristics: Larsen & Toubro Limited (L&T) Petrol Dispensing Pumps & Systems make Fuel Dispensing Pump SPRINT series /Z-line series/PaceMaker-IV series comprising of one or more system in every enclosure and each sub system consist of the following.

1. (L&T) Flow meter types A or B
2. (L&T) Pump and gas eliminator type A or type CPU
3. (L&T) Calculator – Electronic Register Assembly ERA
4. (L&T) Measuring transducer (pulser assembly)

Accuracy Class	Maximum flowrate l/min	Minimum flowrate l/min	Minimum measured quantity 1	Temperature range °C	Maximum unit price (number of digits)	Maximum price to pay (number of digits)	Liquids
1	2	3	4	5	6	7	8
0.5	50/80	4	1	-40 to +50	999999 (Floating Decimal)*	9999999 (Floating Decimal)*	Petrol Diesel Kerosene Petrol mixed with max 15 % Ethanol, 15 % MTBE & 15% Methanol

* - Activation Optional

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.