

OIML Member State
The Netherlands

Number R117/2007-NL1-17.02
Project number 16200138
Page 1 of 2

Issuing authority
Person responsible: NMi Certin B.V.
C. Oosterman

Applicant and
Manufacturer Chengdu Huaqi Houpu Holding CO., LTD
No.555,Kanglong Road, Gaoxin District
Chengdu, Sichuan
China

Identification of the
certified type **A Electronic calculator and indicating device**
Type: HQHP-JQJ-CS

Characteristics See page 2 and further

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 117-1 (2007) "Dynamic measuring systems for liquids other than water"

Accuracy class 0,5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
22 March 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R117/2007-NL1-17.02
Project number 16200138
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated report No. NMI-16200138-01 dated 23 February 2017 that includes 41 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the Electronic calculator and indicating device are presented. The construction of the measuring instrument is recorded in the Documentation folder no. TC8867-1.

Table 1 General characteristics

Environmental classes	M1 / E2
Ambient temperature range	-25 °C – +55 °C; condensing humidity
Intended for the measurement of	liquid
Power supply voltage	230 V AC; 50 Hz
Software identification	Version number: LNGCS 3.066TY. Checksum: 79D6.