

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R129/2000-DE1-08.02

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Ahmed Abou-Zeid

Applicant

Name: TeleFrank GmbH
Address: Am Wildengrund 1, 98553 Altendambach, Germany

Manufacturer of the certified type is the applicant.

Identification of the certified type

Multi-dimensional measuring instrument
Type: LBHG-Modul
Further characteristics: see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R129, Edition 2000 (E)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R129/2000-DE1-08.02

The conformity was established by the results of tests and examinations provided in the associated Report:

No. PTB-5.45-4024517 that includes 8 pages
and associated Test Reports
No. R129/2000-DE-08.02 that includes 65 pages
No. CE-LAB, Pb EMV 030108_01 TeleF that includes 32 pages

The Issuing Authority

The OIML Member

Dr. Ahmed Abou-Zeid
Direktor und Professor

Prof. Dr. R. Schwartz
Direktor und Professor

07.08.2008

07.08.2008

Identification of the type (continued)

Measurement ranges, scale intervals and transport speed

Dimensions	Minimum mm	Maximum mm	scale interval <i>d</i> mm	<i>v</i> m/min
Length	20	360	2	30
Width	20	255	2	
Height	5	80	0,5	

The instrument only measures rectangular and opaque objects

Environmental conditions

Temperature range: +5°C up to +40°C
Humidity conditions: Non-condensing humidity
Intended location: Indoor

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.