# Physikalisch-Technische Bundesanstalt



**Braunschweig und Berlin** 

Member State of OIML Germany



OIML Certificate N° R136/2006-DE1-09.01

### OIML CERTIFICATE OF CONFORMITY

**Issuing Authority** 

Name: Address: Physikalisch-Technische Bundesanstalt Bundesallee 100, 38116 Braunschweig

Person responsible:

Dr. Ahmed Abou-Zeid

**Applicant** 

Name:

**GFal** 

Address:

Rudower Chaussee 30, 12489 Berlin

Manufacturer of the certified type is the applicant.

Identification of the

certified type

Instrument for measuring the areas of leathers

Type: LQScan1809

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R136 Part 1, Edition 2004 (E) R136 Part 2, Edition 2006 (E)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

# Physikalisch-Technische Bundesanstalt



OIML Certificate N° R136/2006-DE1-09.01

The conformity was established by the results of tests and examinations provided in the associated Report:

No. PTB-5.45-4042238

that includes 8 pages

and associated Test Reports

-09.01 that includes 49 pages

No. R136/2006-DE1-09.01 No. ProEMV, PLE090702

that includes 31 pages

(concerning EMC)

The Issuing Authority

Q. Sbou- Zeid

The CIML Member

Dr. Ahmed Abou-Zeid Direktor und Professor

30.09.2009

Dr. Roman Schwartz Direktor und Professor

30.09.2009

#### Identification of the certified type (continued)

Measurement ranges and scale interval

Dimensions	Minimum in mm	Maximum in mm	Scale interval d in dm²
Length	100	1800	1
Width	100	900	

The instrument measures only full-flat objects (no holes) with a translucency under 10% of the visible light.

#### **Environmental conditions**

-Temperature range: 10°C up to + 40°C

-Humidity conditions: non-condensing humidity

-Intended location: Indoor

Important note: Apart from the mention of the Certificate's reference number and the name of the

OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either

may be reproduced in full.