

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R137/2012-NL1-15.03 Project number SO14204809 Page 1 of 3

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and

certified type

Metreg Technologies GmbH

manufacturer Tränkeweg 9

15517 Fürstenwalde

Germany

Identification of the

A rotary displacement gas meter

Type: MRM

Characteristics

See page 2 and 3

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 137-1 (2012) "Gas meters"

Accuracy class

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

16 April 2015

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member StateThe Netherlands

Number R137/2012-NL1-15.03 Project number SO14204809 Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report:

- No. NMi-SO14204809-02 dated 7 April 2015 that includes 12 pages.

Characteristics of the gas meter:

Table 1 gives the general characteristics of the meter type. Table 2 and 3 specify in detail the essential characteristics and verification scale interval.

+ + + + + + + + + + Table 1: Ge	+ + + + + + + + + + Table 1: General characteristics						
Destined for the measurement of	Gas volume						
Mechanical class + + + + + + + +	+ + + + + + + + M2+ + + + + + + + +						
Electromagnetic class + + + + + + +	Not applicable (the meter has no electronics)						
Ambient temperature range	* * * * * * * * -25 °C / +55 °C * * * * * * * *						
Gas temperature range	-25 °C / +55 °C						
Designed for humidity conditions	Not applicable (the meter has no electronics)						
Orientation + + + + + + + + + + + + + + + + + + +	Horizontal, vertical up and vertical down (all orientations)						
Flow direction + + + + + + + + +	+ Uni-directional (indicated with arrow) + + +						
Power supply voltage * * * * * * * *	+ + + + + + Not applicable + + + + + + +						
Software identification	Not applicable						

5



OIML Certificate of Conformity

OIML Member State The Netherlands

Number R137/2012-NL1-15.03 Project number SO14204809 Page 3 of 3

+	Table 2: Essential characteristics					
+	+ Nominal + -	+ +Type+ +	+ + Cyclic + +	+ +Qmax + -	+ + Qt + +	+ + Qmin + +
+	diameter diameter	+ + + + +	volume	+ + + + + +	+ + + + +	+ + + + + +
+	[mm]	+ + + + +	[dm³]	[m³/h]	[m³/h]	[m³/h]
+	+ + 25+ + +	+ +G10 + +	+ + 0,177 + +	+ + 16+ + +	+ + 1,6 + +	+ + 0,4 + +
+	50	G16	0,210	25	2,5	0,5
+	+ + + + + +	G25	0,283	40	2,0	0,5
+		+ +G40 + +	0,566	+ + 65 + + -	3,25	0,5
+	+ + + + + +	+ +G65 + +	+ + 0,708+ +	+ + 100+ + +	+ + 5,0 + +	+ + +0,5 + +
+	80	G100	1,05	160	8,0	0,65
+		G160-3"	2,78	250	12,5	1,6
+	+ + 100+ + +	• G160-4 " •	+ + 2,78 + +	+ + 250+ + +	+ +12,5 + +	+ + +1,6 + +
+	+ + + + + +	+ G250+	+ + 4,20 + +	+ + 400 + +	+ +20,0 + +	+ + +2,0 + +
+	+ + + + + +	G400-4"	† † 5,66 † †	650	* *32,5 * *	+ + 3,2 + +
+ [+]	150	G400-6"	10,5	650	32,5	6,5
+	+ + + + + +	G650	15,7	1000	+ +50,0 + +	10,0
+	+ + 200+ + +	† G1000 [†] †	+ + 19,7 + +	1600	80,0	16,0 + +

Remarks regarding table 2:

- + 1. The overload flow rate (Q_r) for all rotary meters is equal to 1,2 \cdot Q_{max} . + +
- 2. The working pressure range for all rotary displacement gas meters is atmospheric up to and including 16 bar(g).

	Table 3: Verification scale interval							
+ + + Type + + +	+ + + + + + number	+ + control-element + +						
+ + + + + + + +	before the comma	behind the comma	[m3]					
G10 – G25	+ + + + 6 + + + +	+ + + +2+ + + +	0,002					
G40 - G400	+ + + + 7 + + + +	+ + + +1+ + + +	+ + + + 0,02+ + + +					
G650 - G1000	+ + + + + + + + + +	+ + + + 0 + + + +	0,2					

Installation conditions:

For this rotary meter are no specific installation conditions applicable.