

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R137/2012-NL1-15.05 Project number 14200731 Page 1 of 3

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and manufacturer

GFO Europe B.V. Magnesiumstraat 14 6031 RV Nederweert The Netherlands

Identification of the

A rotary displacement gas meter

certified type

Type: GFO-RM

Characteristics

See page 2 and 3

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 137-1 (2012) "Gas meters"

Accuracy class

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

10 June 2015

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R137/2012-NL1-15.05 Project number 14200731 Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMi-14200731-02 dated 4 June 2015 that includes 19 pages.
- No. NMi-SO14200462-02 dated 20 February 2014 that includes 13 pages.
- No. NMi-12200078-01 dated 29 May 2012 that includes 51 pages.
- No. NMi-10200626-02 dated 12 October 2011 that includes 50 pages.

+ Characteristics of the gas meter:

Table 1 gives the general characteristics of the meter type. Table 2 and 3 specify in detail the essentia characteristics and verification scale interval.

Table 1: General characteristics					
Destined for the measurement of	+ + + + + + Gas volume				
Mechanical class	+ + + + + + + + M2				
Electromagnetic class	Not applicable (the meter has no electronics)				
Ambient temperature range	-25 °C / +55 °C				
Gas temperature range + + + + + + +	+ + + + + + -25 °C / +55 °C + + + + + + +				
Designed for humidity conditions	Not applicable (the meter has no electronics)				
Orientation	Horizontal, vertical up and vertical down (all orientations)				
Flow direction	Uni-directional (indicated with arrow)				
Power supply voltage	Not applicable				
Software identification	Not applicable				

5



OIML Certificate of Conformity

OIML Member State The Netherlands

Number R137/2012-NL1-15.05 Project number 14200731 Page 3 of 3

Table 2: Essential characteristics						
+ + Cyclic + +	+ +Type+ +	+ + Qmax + +	+ + Qt+ + +	Qmin	Nominal	
volume	+ + + + +	+ + + + + +	+ + + + + +	+ + + + +	diameter	
[dm³]		[m³/h]	[m³/h]	[m³/h]	[mm]	
	G10	16	1,25	0,5	40	
0,26	G16	25	1,25	0,5	40	
+ + + + + +	G25	40	2	0,5	40	
	G16	25	1,25	0,5	40 / 50	
0.60	G25	40	+ + + + + + + + + + + + + + + + + + + +	0,5	40 / 50	
+ + 0,69	+ + G40 + +	65	3,25	0,5	40 / 50	
+ + + + + +	+ +G65 + +	+ + 100 + +	+ + +5 + + +	+ + 0,5 + +	+ + 40 / 50 + +	
+ + + + + +	+ +G40 + +	+ + +65+ + +	+ + 3,25+ + +	+ + 8,0 + +	+ + 50 / 80 + +	
+ + 1,11+ + +	+ +G65 + +	+ + 100 + +	+ + + 5 + + +	+ + 0,8 + +	+ 50 / 80 + +	
	G100	160	8 + 4	0,8	50 / 80	
	G65	100	12,5	1,25	80 / 100	
2,31	G100	160	12,5	1,25	80 / 100	
	G160	250	12,5	1,25	80 / 100	
* * * * * *	G100	160	8	2 * *	80 / 100	
2,98	G160	250	12,5	+ + 2 + +	80 / 100	
+ + + + + +	G250	400	+ + 20+ + -	+ + 2 + +	+ +80 / 100 +	
3,88 1)	+ G250S+ +	+ + 400 + +	+ + 20+ + +	+ +3,25 + +	+100 / 150	
+ + + + + + +	+ G400S+ +	+ + 650 + +	+ + 32,5+ + +	+ +3,25 + +	+ +100 / 150 +	
+ +5,97 ¹⁾ + +	G400S	+ + 650 + +	+ + 32,5	+ + 5 + +	+ + 150 + +	
7,81	G650S	1000	50	+ + 5 + +	150	

Remarks regarding table 2:

- 1. Twin rotor version named "G...S".
- 2. The overload flow rate (Q_r) for all rotary meters is equal to $1.2 \cdot Q_{max}$.
- 3. The working pressure range for all rotary displacement gas meters is atmospheric up to and including 20 bar(g).

Table 3: Verification scale interval							
+ + + Type + + +	+ + + + + + number	control-element					
	before the comma	behind the comma	[m3]				
G10 – G65	6	2	0,002				
G100 – G650	7	1	0,02				

Installation conditions:

The meter can be installed in horizontal, vertical up and vertical down position. Regarding flow disturbance there are no specific installation requirements.