



NMI Australia  
Legal Metrology Policy  
Att. Dr. Phillip Mitchell  
[Phillip.Mitchell@measurement.gov.au](mailto:Phillip.Mitchell@measurement.gov.au)

Document for comments on Committee Draft			NL
Comments on: <b>OIML TC 12</b>	Draft Recommendation: OIML 1DR R 46-3	Title: Active Energy Electricity Meters; Part 3: Test report format	Project: <b>p1</b> ; Revision of OIML R 46
1DR date: <b>10 January 2013</b>	Circulation date: <b>10 January 2013</b>	<b>Closing date for comments:</b> <b>12 April 2013</b>	
Secretariat: <b>Australia</b> Dr. Phillip Mitchell	<b>Please include any comments in this template</b> and send it <b>in word</b> format as soon as possible and <b>→ not later than the closing date ←</b> preferably <b>by e-mail</b> to the secretariat (see above address)		

Country Code	Clause/ paragraph/ table	gen./ edit./ techn.	COMMENTS	PROPOSED CHANGE	
NL	6.17 Dry Heat	gen.	Need for check any register change as consequence of the exposure to dry heat	Suggest to add: Check any change of register < critical change value	
NL	6.18 Cold	gen.	Need for check any register change as consequence of the exposure to cold	Suggest to add: Check any change of register < critical change value	